Hospital Bed





Specification:

* Part

1. Actuator Mechanism

Actuator can be adjust by control that are easily reahable by patient.

2. Head & foot panel

Head & foot board made of HDPE, detachable & easy to clean. Equipped with bumper (with humanization design for anti collision effect), located at each corner

3. Wheels

Four wheels 5" double wheel, equipped with brakes on two of them

4. Side quard

Made of aluminium alloy, with collapsible system and can be foled

5. Infuse stand

Made of stainless steel pipe, adjustable height with screw lock. Two hanger (hook).

6. Mattress

Dimension 1900 (L) x 860 (W) x 100 (H) mm

7. Urine Hanger

Made of metal & Powder Coating Finishing





* Position

Back raise lifting angle : 0 - 60° (Approx) Knee raise lifting angle : 0 - 30° (Approx)

* Dimension (Approx)

Overall 2140 (L) x 970 (W) x 970 (H) mm

* Attention !



Maximum Load 200 Kg

* Direction

Push the triangle with upwards point to rise the back and knee. push the triangle downward point to lower.

* TKDN

Nilai: 46.87%

Sertifikat: 784/SJ-IND.8/TKDN/1/2023

* Scan here to visit webpage



